

Flow Measurement Specialists

Systems | Sonic Nozzles | Venturis | ASME Flow Nozzles | Orifice Meters | Consulting





THE OPERATIONS MANUAL DETAILS SAFE OPERATING PROCEDURES AND RECOMMENDED SAFETY EQUIPMENT FOR THE OPERATOR.

TO AVOID POSSIBLE PERSONAL INJURY AND DAMAGE TO EQUIPMENT; COMPLETELY FAMILIARIZE YOURSELF WITH THE CONTENTS OF THIS OPERATIONS MANUAL BEFORE TURNING ON POWER OR OPERATING THIS EQUIPMENT.

ACAUTION



wear eye protection





wear ear protection



FOR OPERATOR SAFTEY, HEARING AND EYE PROTECTION MUST BE USED WHILE OPERATING THIS TEST STATION.

THE SOUND LEVEL EMITTED DURING TESTING CAN EXCEED 122 DBA. BECAUSE THE NOISE IS GENERATED BY THE PART UNDER TEST, IT IS THE CUSTOMERS RESPONSIBILITY TO MAKE SOUND PRESSURE LEVEL MEASUREMENTS AND TAKE NECESSARY ACTIONS TO PROTECT OPERATORS AND BY-STANDERS FROM NOISE RELATED INJURIES AND TO COMPLY WITH ANY LOCAL OR COMPANY REGULATIONS

220 Bunyan Avenue Berthoud, CO 80513 Phone 970.532.0617 Fax 970.532.0748 www.flowsystemsinc.com

Page 1